

**NO.D.12016/2/2016-NHM/DHME/NPHCE  
NATIONAL PROGRAMME FOR HEALTH CARE OF THE ELDERLY  
NATIONAL HEALTH MISSION  
HEALTH & FAMILY WELFARE DEPARTMENT**

*Dated Aizawl, the 10<sup>th</sup> July, 2013.*

**TENDER NOTICE**

Sealed Tenders are hereby invited on behalf of the Chairman, Mizoram State Health Society from reputed and bonafide Manufacturers or their authorized dealers who have Mizoram VAT Registration for **supply of Hospital Furniture & Appliances etc for District Hospitals in Mizoram under NPHCE** which will be received by the Mission Director, National Health Mission, Mizoram on or before **31.8.2013, upto 12:00 PM** under the terms and conditions laid down below. Quotations are to be opened by MD NHM, Mizoram, Aizawl or his authorized representatives at **1:00 PM** on the same day. The Tenderers or their representative may also be present at the time of opening of the Quotations, if they so desire.

**TERMS & CONDITIONS**

1. Tenders submitted by Authorized dealers shall attach Authorization or Distributorship or Dealership (as the case may be) certificates from the concerned Manufacturers. A Gazetted Officer or Notary, however, shall duly attest Photo copy. Original Certificate shall be produced at the time of opening, if demanded.
2. The rate should be quoted as per the technical specifications to be offered by the Mizoram state Health society against the equipments to be quoted.
3. The technical and Financial Bid shall be submitted separately.
4. Tenders shall be submitted in the prescribed form only. A soft copy of prescribed form can be obtained from State NCD Cell, H&FW Dept. and shall be submitted along with the hard copy.
5. Earnest Money amounting to Rs. 10,000/- (Rupees ten thousand) only in the form of Bank Draft, from any Nationalized Bank of India or Postal Saving Certificate duly pledged in favour of Mission Director, National Health Mission, Mizoram, Aizawl must accompany such tender.
6. The cover of the tenders shall be superscribed as **“supply of Hospital Furniture & Appliances etc for District Hospitals in Mizoram”** addressed to Mission Director, National Health Mission, Mizoram, Aizawl clearly indicating full postal address of the tenderer/sender.
7. Non-tribal tenderer shall affix a Non- Judicial Court Fee Stamp worth at least Rs.7.50/- and shall furnish photo copy of MVAT Registration Certificate and MVAT Clearance Certificate up to the latest year for which assessment has been made by the concerned authority and duly attested by a Gazetted Officer or Notary. Original Certificates shall be produced at the time of opening, if demanded.
8. The tribal tenderer shall furnish attested copy of Tribal Certificate, MVAT Registration Certificate and MVAT Clearance Certificate up to the latest year for which assessment has been made by the concerned authority and duly attested by a Gazetted Officer or Notary. Original Certificates shall be produced at the time of opening, if demanded.

9. Manufacturers submitting their tenders directly shall attach photo copies of their Manufacturing Licence, CIN (Corporate Identity Number) and other relevant permissions/licences required for such manufacturing, duly attested by a Gazetted Officer or Notary. Original Certificates shall be produced at the time of opening of the tenders, if demanded.
10. Tenderer shall submit Notarized Photo copy of valid and up to date GMP/ISO Certificate.
11. Tenderer shall quote their rates in terms of Indian Rupees.
12. Rate shall be quoted inclusive of all taxes and duties and also clearly mentioned in the quotation.
13. F.O.R Destination shall clearly be mentioned.
14. All pages of the tender papers shall be printed, signed, stamped and dated; hand written quotations shall not be entertained. Any page or pages not signed, stamped and dated are liable to be rejected. Correction, if any, shall be initialed, stamped and dated.
15. If any tenderer violates or fails to fulfill any one of the clauses, his or her tender are liable to be rejected at the time of opening the tender.
16. Once the rate is accepted and communicated to the successful firm, the same will remain valid during the validity period of the agreement with the approved supplier and no price escalation will be permissible during the validity period, which is normally for a period of 12 months and is extendable, if mutually agreed upon by both the parties i.e. the Health & Family Welfare Department and concerned approved supplier, subject to recommendation to the Purchase Committee of Mizoram State Health Society.
17. A successful tenderer whose tender has been accepted shall execute an appropriate Deed of Agreement within 15 (fifteen) days from the date of communication of acceptance of his/her quoted rate(s) by fax, failing which their tender will be treated as invalid.
18. A successful firm shall supply the materials within the period fixed by the Department/Authority. If the supply cannot be accomplished within the stipulated period without any intimation, the Department has the right to procure the materials from any other sources and the differences in prices, if any, will be realized from the Security Deposit of the approved supplier, as mentioned herein, without assigning any reason thereof.
19. A successful firm will have to furnish Security Deposit of 5% of the value of the supply before issue of supply order in the form of Demand Draft/Fixed Deposit in any Nationalized Bank of India or Postal Saving Certificate, as the case may be, duly pledged in favor of Mission Director, National Health Mission, Mizoram
20. The State Govt. shall not be bound to accept the lowest rate or any of the tenders even if valid, and reserves the right to accept or reject any tender without assigning any reason thereof. The Govt. shall be at liberty to prefer the quoted rates for materials/equipments of reputed brand/company, with ISO specifications compared to other rates. Tenderers having satisfactory performance in the supply, installation, servicing and maintenance of such medical equipments within the state of Mizoram may enclosed Supporting documents.

21. In case of any legal dispute/litigations arising out of this Tender Notice or its terms, the local courts shall have the jurisdiction as per the relevant laws in force.
22. Payments will be made only after the materials are received in full and in good condition to be certified by the Competent authority and after completion of the installation satisfactorily at the representative destination whatever it is required so. There shall not be any advance payment and no such request shall be entertained.
23. Any attempt for negotiation/campaign, direct or indirect, on the part of a tenderer with the authority to whom he has submitted the Tender or the authority who is competent to accept the same, shall render such Tender liable to be excluded from consideration, if it is so proved.
24. Belated/delayed tenders shall not be entertained in any case. Certificates other than those stated in the foregoing clauses shall be rejected straightaway. Any other/further clarifications in connection with submission of tender mentioned herein may be made from the office of Mission Director, National Health Mission, Mizoram, Aizawl during office hours on all working days.
25. A successful tenderer shall have to supply the intended equipment to Central Medical Store, Zemabawk at his own risk and responsibility and in good condition all respects.
26. Incomplete tender not in conformity with the terms and conditions as in this tender notice will be rejected outright.
27. Decision of the Purchase Committee of the Society shall be final and binding in all cases.
28. A successful tenderer should clearly mentioned the warranty period offered.

Sd/-  
**( Dr.K. LALBIAKZUALA )**  
Secretary cum Mission Director,  
National Health Mission,  
Mizoram State Health Society.  
Aizawl

Memo No.D.12016/2/2016-NHM/DHME/NPHCE : Dated Aizawl, the 10<sup>th</sup> August,2016.

Copy to :

1. P.S to Minister, H & F W Deptt., for favour of information.
2. P.S to Secretary, H & F W Deptt., for favour of information.
3. The Principal Director, H & F W Deptt.,
4. The Director, DHME., Mizoram for kind information
5. The Director, DHS., Mizoram for kind information

Secretary cum Mission Director,  
National Health Mission,  
Mizoram State Health Society.  
Aizawl

Annexure

<b><u>TENDER FOR SUPPLY OF HOSPITAL FURNITURE &amp; APPLIANCES</u></b>							
<b>National Programme for Health Care of the Elderly (NPHCE)</b>							
<b>Sl.No</b>	<b>Items</b>	<b>Specifications/ Description</b>	<b>Qty</b>	<b>Unit Rate in figure</b>	<b>Unit Rate in word</b>	<b>Total Amount in Figure</b>	<b>Total Amount in Word</b>
1	Fowler bed	1. Steel chrome finished, 2. Head bow & foot bow, 3. Fitted with side railing & wheel with lock, 4. Back rest & knee rest fitted with separate adjustable screw from foot end, 5. ISI/ISO/EC or equivalent certified	20				
2	Ward bedside locker	1. Metal cupboard underneath with stainless steel plated, knob handle 2. 16 inch x 14 inch x 32 inch 3. ISI/ISO/EC or equivalent certified	20				
3	Revolving Stool	1. Steel tube legs with adjustable screw to raise or lower the top 2. Height adjustment from 18" to 27" with stainless steel top 12" diameter fitted with rubber shoes white enamel stove heated.	20				
4	Drip stand	1. Tubular frame fitted on steel or plastic heavy base with ball bearing castors 2. Stainless steel chrome finished.	20				
5	Ward Screen Four Folds	1. Tubular steel frame mounted on eight ball bearing castors provided with hooks and wire springs for curtains and complete with curtain cloth 2. White enamel stove painted 3. Each fold measures 5ft.6" X 1 ft.10"	10				
6	Oxygen Cylinder Trolley	Tubular steel frame to hold oxygen cylinder provided with castors white enamel stove heated.	10				
7	Hand Stretcher Folding Type	Wooden handle and poles metal cross bars and canvas top	10				
8	Examination Table	<ul style="list-style-type: none"> <li>• Stainless Steel Frame 72" long, 20" wide, 32" high</li> <li>• 16 G Sheet Steel top on angle frame.</li> <li>• Adjustable Headpiece fitted with foam mattress.</li> <li>• Legs fitted with wheels with lock.</li> <li>• It should be ISI/ISO/EC or equivalent certified</li> </ul>	4				

9	OPD & Ward tables	Godrej T-32 model	6				
10	Chairs	Revolving chairs, PHC-7202RX	6				
11	Almirah	Storewell big plain	3				
10	Mattress	6.5 x 3 ft	20				
12	Blankets	6x3 ft * light weight * Machine washable	60				
13	Bed sheets	* Size 65" X 100" * Smooth texture & Durability * Light weight & Comfortable * Available for Single Hospital Bed * White bedsheet cotton	60				
14	Bed cover	* suitable for single Hospital Bed * Wrinkle resistance * Skin-friendliness * Softness	60				
15	Pillows with cover	1x2 ft * lightweight, comfortable & soft	20				
16	Nebuliser	<ul style="list-style-type: none"> <li>• Non-heating compressor nebulizer</li> <li>• Electrical: 230V, 50Hz model</li> <li>• Nebulizer Rate: 0.4 ml/min (without cap)</li> <li>• Medication Cup Capacity: 7ml (cc) max</li> <li>• Particle Size: *MMD approximately 5 μ</li> <li>• Should be provided with a complete nebulisation kit of 10 Nos. including adult and child mask and medication cup – 5 Nos.</li> </ul>	15				
17	ECG Machine	<ul style="list-style-type: none"> <li>• Simultaneous 3 channel ECG recording with 12 lead simultaneous acquisitions with auto summary</li> <li>• Built in ECG parameters measurement and interpretation</li> <li>• Dimensions 300mm x 260mm x 75m</li> <li>• Display 192 x 64 pixels LCD (Narrow Screen)</li> <li>• Light weight design for portability-2.5 kg.</li> <li>• High performance Li-ion rechargeable battery.</li> <li>• Power line safe design, can be used both on power line as well battery.</li> <li>• Digital filtering remove interference from power line, muscle tremor etc., with automatic base line correction.</li> </ul>	7				

18	Pulse Oximeter	<ul style="list-style-type: none"> <li>• Should have plethysmographic wave form with numeric display for SPO2 and Heart rate on LCD/TFT display</li> <li>• Should have a SPO2 range of 0 to 100%</li> <li>• Sensor for neonates/ infants and paediatric</li> <li>• Should provide bar graph for pulse strength</li> <li>• Beep sound and alarm sound should have separate volume control</li> <li>• Should have a minimum of 2 hours back-up time.</li> <li>• Should be a portable, light weight and desktop model.</li> <li>• Should work with input 200 to 240Vac 50 Hz supply</li> </ul>	15				
19	Weighing machine	<ul style="list-style-type: none"> <li>• Personal Weighing Machine(Capacity 150 kg , Graduation 0.5 kg)</li> </ul>	4				
<b>TOTAL</b>							
				<b>Bidder's Name</b>	: _____ _____		
				<b>Signature</b>	: _____ _____		
				<b>Date</b>	: _____ _____		
				<b>Stamp</b>			